

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 24 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130133 ✓
A	Name	Bryan Barlow	ID# 4548 Date 6/21/09 ✓
	Agency	Alaska State Troopers	Phone # 907-783-0972
	Instrument Location	Girdwood - AST ✓	
B	Alco S/N	75260 ✓	Target Value .079- High Pressure 1150
	Alco Test Value Average	.079 ✓	.079 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature	Bryan Barlow ✓	

(OVER)

BMB 6/25/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:  
(Do Not Write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 20th day of June, 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

JUNE 21, 2009

OPERATOR'S NAME:

BARLOW BRYAN U

OPERATOR'S NUMBER: 4548

SUBJECT'S LAST NAME:

CALIBRATION

SUBJECT'S FIRST NAME/ID:

VERIFICATION OF

O.L. #: 1234567

DEPT/AGENCY: 6040

CASE/REPORT: 09-12345

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: 75268

### — BREATH ANALYSIS —

.079 ADJUSTED FOR 30.16 in ✓  
ALCO TARGET .079 22:12  
BLANK TEST .000 22:13  
INTERNAL STANDARD VERIFIED 22:13  
ALCO TV 30.16 in .079 ✓ 22:14  
BLANK TEST .000 22:15  
SUBJECT SAMPLE .000 22:15  
BLANK TEST .000 22:16  
ALCO TV 30.16 in .079 ✓ 22:17  
BLANK TEST .000 22:17

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

JUNE 21, 2009

TIME 22:09

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.16 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~!@